



Phones : (0522) 2668004-008, 700, 800, 900
Fax : (0522) 2668017, 2668129

Department of Microbiology

संजय गाँधी स्नातकोत्तर आयुर्विज्ञान संस्थान, लखनऊ
Sanjay Gandhi Postgraduate Institute of Medical Sciences

Raebareli Road
Lucknow-226 014 (India)

Date: 09/11/2015

Microbiocidal Testing Report

This report is to state that OptiMaser Microwave Medical Waste Disinfection System (M.M.D.S.) has good microbiocidal efficacy. In one cycle run of forty minutes, 105 organisms of ATCC *Escherichia coli* 25922 were killed yielding zero colony count.

Dr. T.N. Dhole
Prof. & Head
Dept. of Microbiology
SGPGIMS, Lucknow

Prof. T. N. DHOLE
Prof. & Head
Dept. of Microbiology
S.G.P.G.I.M.S., Lucknow